



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

EIZO Corporation
% Hiroaki Hashimoto
Manager
153 Shimokashiwano
Hakusan, Ishikawa 924-8566
JAPAN

April 3, 2015

Re: K150106
Trade/Device Name: 2MP Color LCD Monitor, RadiForce RS240
Regulation Number: 21 CFR 892.2050
Regulation Name: Picture archiving and communications system
Regulatory Class: II
Product Code: PGY
Dated: January 9, 2015
Received: January 20, 2015

Dear Hiroaki Hashimoto:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, “Misbranding by reference to premarket notification” (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH’s Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink that reads "Robert A. Ochs". The signature is written in a cursive style. Behind the signature, there is a faint, large, light-gray watermark of the letters "FDA".

Robert Ochs, Ph.D.
Acting Director
Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K150106

Device Name

2MP Color LCD Monitor, RadiForce RS240

Indications for Use (Describe)

This product is intended to be used in displaying and viewing digital images for review, analysis and diagnosis by trained medical practitioners. It does not support the display of mammography images for diagnosis.

Type of Use (Select one or both, as applicable)

☒ Prescription Use (Part 21 CFR 801 Subpart D)

☐ Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

“An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number.”



EIZO Corporation, 153 Shimokashiwano, Hakusan, Ishikawa
924-8566 Japan

U.S. Food and Drug Administration
Center for Devices and Radiological Health
Document Mail Center – WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

Name Hiroaki Hashimoto
Department Medical System Standards

Telephone +81 (76) 274-2468
Fax +81 (76) 274-2484
E-Mail hiroaki.hashimoto@eizo.com

510(k) Summary (in accordance with 21 CFR 807.92)

1. Company

EIZO Corporation
153 Shimokashiwano, Hakusan
Ishikawa 924-8566 Japan
Tel: +81 (76) 274-2468
Fax: +81 (76) 274-2484

2. Contact Person

Hiroaki Hashimoto

3. Date of Summary

January 16th, 2015

4. Device Information

- Trade Name/Model: RadiForce RS240
- Common Name: 2MP Color LCD Monitor
- Classification Name: Display, Diagnostic Radiology
- Regulation Number: 21 CFR 892.2050, Product Code PGY

5. Predicate Device

- 2MP Color LCD Monitor, RadiForce RX240 (K113844)

6. Device Description

RadiForce RS240 is a color LCD monitor for viewing medical images other than those of mammography. The color panel employs in-plane switching (IPS) technology allowing wide viewing angles and the matrix size (or resolution) is 1,200 x 1,600 pixels (2MP).

Since factory calibrated display modes, each of which is characterized by a specific tone curve (including DICOM GSDF), a specific luminance range and a specific color temperature, are stored in lookup tables within the monitor, the tone curve is e.g. DICOM compliant regardless of the display controller used.

RadiCS is application software to be installed in each workstation offering worry-free quality control of the diagnostic monitors including the RadiForce RS240 based on the QC standards and guidelines and is capable of quantitative tests and visual tests defined by them. The RadiCS and its subset, RadiCS LE, are included in this 510(k) submission as an accessory to the RadiForce RS240.

7. Intended Use

This product is intended to be used in displaying and viewing digital images for review, analysis and diagnosis by trained medical practitioners. It does not support the display of mammography images for diagnosis.

8. Comparison of Technological Characteristics

The comparison table below enumerates information derived from the product brochure of the each device and different technological characteristics are discussed in it:

Attributes	RadiForce RS240	RadiForce RX240	Explanation of Differences
Display Performance/Specifications			
Screen technology	IPS TFT Color LCD Panel	IPS TFT Color LCD Panel	-
Viewing angle (H, V)	H: 176°, V: 176°	H: 176°, V: 176°	Typical data for very low contrast provided by the panel manufacturer is cited.
Resolution	2MP (1,200 x 1,600)	2MP (1,200 x 1,600)	-
Aspect ratio	3 : 4	3 : 4	
Active screen size	324.0 mm x 432.0 mm	324.0 mm x 432.0 mm	
Pixel pitch	0.270 mm x 0.270 mm	0.270 mm x 0.270 mm	-
Maximum luminance	800 cd/m ²	760 cd/m ²	-
DICOM calibrated luminance	400 cd/m ²	400 cd/m ²	-
Contrast ratio	1400 : 1	1200 : 1	Typical contrast ratio data provided by panel manufacturers is cited.
Backlighting	LED	LED	-
Display Colors	From a palette of 68 billion colors: - 10-bit input (DisplayPort): 1.07 billion colors (maximum) - 8-bit input: 16.77 million colors	From a palette of 68 billion colors: - 10-bit input (DisplayPort): 1.07 billion colors (maximum) - 8-bit input: 16.77 million colors	-
Luminance non-uniformity compensation	Digital Uniformity Equalizer	Digital Uniformity Equalizer	-

Video Signal Input			
Input video signals	DVI-D x 1, DisplayPort x 1	DVI-D x 1, DisplayPort x 1	-
Scanning Frequency (H, V)	31 - 100 kHz / 59 – 61 Hz (VGA Text: 69 - 71 Hz) Frame synchronous mode: 59 - 61 Hz	31 - 100 kHz / 59 – 61 Hz (VGA Text: 69 - 71 Hz) Frame synchronous mode: 59 - 61 Hz	-
Power Related Specifications			
Power Requirements	AC 100 - 240 V: 50 / 60 Hz	AC 100 - 120 V, 200 - 240 V: 50 / 60 Hz	-
Power Consumption / Save Mode	79 W / Less than 1.6 W	105 W / Less than 1.6 W	-
Power Management	DVI DMPM, DisplayPort 1.1a	DVI DMPM, DisplayPort 1.1a	-
Miscellaneous Features/Specifications			
QC software	RadiCS	RadiCS	-
Sensors	Backlight Sensor, Presence Sensor	Backlight Sensor, Presence Sensor, Integrated Front Sensor (IFS), Ambient Light Sensor	The both devices are capable of QC tests and calibration with an external sensor.
USB Ports / Standard	1 upstream, 2 downstream / Rev. 2.0	1 upstream, 2 downstream / Rev. 2.0	-
Dimensions w/o stand (W x H x D)	376 x 505 x 98 mm	376 x 505 x 98 mm	-

It is clear that the technological characteristics differences discussed above do not affect the safety and the effectiveness of the RS240.

9. Performance Testing

The following bench tests were performed on the RadiForce RS240.

- Verification of the conformance to DICOM GSDF as specified in *Assessment of Display Performance for Medical Imaging Systems* by AAPM Task Group 18 (TG18 guideline)
- Measurement of the luminance non-uniformity characteristics of the display screen as specified in the TG18 guideline
- Measurement of the chromaticity non-uniformity characteristics of the display screen as specified in the TG18 guideline
- Measurement of the chromaticity at the center of the display screen at 5%, 50% and 95% of the maximum luminance as specified in *Guidance for Industry and FDA Staff: Display Accessories for Full-Field Digital Mammography Systems-Premarket Notification (510(k)) Submissions*
- Visual check of presence or absence of miscellaneous artifacts on the display screen as specified in the TG18 guideline
- Measurement of spatial resolution expressed as modulation transfer function (MTF)
- The maximum number allowed for each type of pixel defects/faults

The test results showed that the RadiForce RS240 has display characteristics equivalent to those of the predicate device, RadiForce RX240 except some items, each of which was determined that it would not affect observer's performance.

Besides, the display characteristics of the RadiForce RS240 meet the pre-defined criteria when criteria are set.

No animal or clinical testing was performed on the RadiForce RS240.

10. Conclusion

The RadiForce RS240 was determined to be substantially equivalent to the predicate device due to the following reasons:

- The stated intended use is substantially the same as that of the predicate device.
- It was confirmed that the technological characteristics different from those of the predicate device do not affect the safety and the effectiveness.
- The bench tests demonstrated that the display characteristics are equivalent to those of the predicate device except some items, each of which was determined that it would not affect observer's performance.